



QPS Evaluation Services Inc.
Testing, Certification and Field Evaluation Body
Accredited in Canada, the USA, and Internationally

File
ML1621

CERTIFICATE OF COMPLIANCE
(ISO TYPE 3 CERTIFICATION SYSTEM)
CERTIFICATION

Issued to	ION SCIENCE LTD.
Address	The Hive, Butts Lane, Fowlmere SG8 7SL, United Kingdom
Project Number	ML1621-1
Product	Portable Gas Detector
Model Number	ARA100, ARA200, ARA300, ARA400, ARA100H and ARA200H
Trademark	ION SCIENCE
Ratings	Class I, Division 1, Groups A, B, C and D, T4; Class I, Zone 0, AEx ia IIC T4 Ga; Ex ia IIC T4 Ga; -40 °C ≤ Ta ≤ +60 °C: ARA100, ARA200, ARA400, ARA100H, ARA200H -30 °C ≤ Ta ≤ +60 °C: ARA300 3.6V battery powered
Applicable Standards	CAN/CSA C22.2 No. 60079-0:19 CAN/CSA C22.2 No. 60079-11:14 ANSI/UL 60079-0 7th ed. ANSI/UL 60079-11 6th ed.

Statement of Compliance: The product(s) identified in this Certificate and described in the Report covered under the above referenced project number have been investigated and found to be in compliance with the relevant requirements of the above referenced standard(s). As such, they are eligible to bear the QPS Certification Mark shown below, in accordance with the provisions of QPS's Service Agreement.



Issued by: Alfonso Mattucci - QPS Quality Assurance

Signature: *Alfonso Mattucci*

Date: July 27, 2020