



No. 2015 - 0423

## Safety Certificate

- Manufacturer **Ion Science limited**
- Address **A the Way Fowlmere Cambridgeshire SG8 7UJ**

The following product fabricated by the above manufacturer has been certified for the use of the Safety Certification Mark as the product was assessed and tested in accordance with the article 34 of the Occupational Safety & Health Act and the article 58 (4) 4 of the Enforcement Rule thereof.

- Product **Phocheck Tiger**
- Type(Model) **Tiger**
- Rating **Um = AC 42.4 V, DC 60 V, Uo = 6 V**
- Type of Protection **Ex ia IIC T4**
- Certification No. **15-GA4BO-0423**
- Safety Certification Standards **article 34 of the Occupational Safety & Health Act**
- Certification condition **Tamb. = -15°C to +45°C (with Li-Ion Battery pack)  
Tamb. = -15°C to +40°C (with Alkaline Battery pack)**
- Date of issue **June 10th, 2015**

*Approved for issue on behalf of the president :* **Gi-Hoi, Kim**  
*position :* **General Manager**

Signature





**Korea Gas Safety Corporation**